

Appendix A

DF or DP ERRORS

- Check for the following:
1. Level detector connection to back of analyzer are tight.
 2. Leakage under instrument.
 3. Level detector inserted correctly into container.
 4. No pinch or kink in reagent tubing.
 5. Check Diluent container level.
 6. Check that waste tube is not pinched or kinked.

All highlighted areas of flowchart are to be performed by trained in-house technician ONLY.

